

## DECLARATION OF CONFORMITY

### Seametrics Inc.

Certifies that the following rate and total indicator products conform with the protection requirements of Council Directive 2014/30/EU and the RoHS requirements of Council Directive (EU) 2015/863:

FT430, FT440, FT450

#### EMC standards applied:

EN61326-1:2013 Electrical Equipment for Measurement, Control and Laboratory EMC Requirements; General Requirements (Industrial)

EN61326-2-3:2013 Electrical Equipment for Measurement, Control and Laboratory EMC Requirements; Particular Requirements, Test Configurations, operational conditions and performance criteria for transducers with integrated or remote signal conditioning

#### Test Methods:

*Radiated and Conducted Emissions:*

CISPR11:2009 (amended by A1:2010)

*Immunity (EN 61326-1:2013) Test Methods:*

IEC61000-4-2:2008 ESD

IEC61000-4-3:2010 Radiated Immunity

IEC61000-4-4:2012 EFT

IEC61000-4-5:2005 Surge


IEC61000-4-6:2013 Conducted Immunity

IEC61000-4-8:2009 Magnetic Field Immunity

IEC61000-4-11:2004 Voltage Interruptions

IEC61000-4-11:2004 Voltage Dips

Issued: October 2019

Signed:  \_\_\_\_\_

Fred Sutton Engineering Manager